



science for animal care

CERTIFICATE of ANALYSIS

Product Name : CDV Ag

Lot No. : 20210528.01

Expiration date : 2023/05/27

1. Appearance

Acceptance Standard : Clean, level off

Results of Inspection : ☒ Approved ☐ Not Good

2. Performance

	Test Item	Signal ratio
Negative	Sample#01 (-)	C6, T0
	Sample#07 (-)	C6, T0
	Sample#15 (-)	C6, T0
	Sample#18 (-)	C6, T0
	Sample#22 (-)	C6, T0
	Assay buffer	C6, T0
Positive	CDV onderstepoort, 1:100	C6, T1

3. Liquid Flow Speed (mm/min)

Acceptance Standard : >10

Results of Inspection : 120 ☒ Approved ☐ Not Good

4. Decision Time (min)

Results of Inspection : ☐ Five/5 ☒ Ten/10

Results of Inspection : ☒ Approved ☐ Not Good

Explanation

Quality of test sample represents this batch

Condition : Temperature $30 \pm 10^{\circ}\text{C}$, Relative Humidity $65 \pm 15\%$

Standard card has a range of color density from 0 to 7

Decision : ☒ Approved ☐ Not Good

