



science for animal care

CERTIFICATE of ANALYSIS

Product Name : CPV Ag

Lot No. : 20210208.03

Expiration date : 2023/02/07

1. Appearance

Acceptance Standard : Clean, level off

Results of Inspection : ☒ Approved ☐ Not Good

2. Performance

	Test Item	Signal ratio
Negative	Sample#02 (-)	C6, T0
	Sample#06 (-)	C6, T0
	Sample#09 (-)	C6, T0
	Sample#14 (-)	C6, T0
	Sample#20 (-)	C6, T0
	Assay buffer	C6, T0
Positive	CPV STRAIN 154, 1:20	C6, T1

3. Liquid Flow Speed (mm/min)

Acceptance Standard : >10

Results of Inspection : 120 ☒ Approved ☐ Not Good

4. Decision Time (min)

Results of Inspection : ☐ Five/5 ☒ Ten/10

Results of Inspection : ☒ Approved ☐ Not Good

Explanation

Quality of test sample represents this batch

Condition : Temperature 30±10°C, Relative Humidity 65±15%

Standard card has a range of color density from 0 to 7

Decision : ☒ Approved ☐ Not Good

